Part # 11201400 DWG # 6598 Rev. B



Geotech Environmental Equipment, Inc. 2650 East 40th Avenue Denver, Colorado 80205 1-800-833-7958 website: www.geotechenv.com

GEOSUB 2 PUMP & CONTROLLER INSTRUCTIONS & MAINTENANCE

BASIC OPERATION

- Plug power cord into controller.
- Plug power cord into AC supply outlet.
- Wait for initialization sequence to complete.
- From main menu:
- Press to start the pump at default settings.
- Wait for pump to finish.



BT-START

STARTING

PLEASE WAIT

##mara

:#8-8"\s"\

- Press or to adjust pump speed to achieve desired flow rate.
- Pump water at desired pump speed.
- Press to stop pump and return to main menu.


DISASSEMBLY AND CLEANING

- 1. Disconnect pump cable from controller.
- 2. Unscrew outer housing (No tools needed).
- 3. Remove inner housing by pulling apart. Slightly twisting while pulling may help.
- 4. Disconnect the connector.
- 5. Remove motor assembly from inner housing by pulling apart, slightly twisting while pulling may help.
- 6. Inspect motor assembly, impeller, and control module for damage or wear.
- 7. Inspect internal wires for cuts or breaks, making sure wires have insulation covering them.
- 8. Inspect all O-rings for nicks, cuts, brittleness.
- 9. Decontaminate outer and inner housing.
- 10. Decontaminate the motor assembly and control module.
- 11. Decontaminate top cap and hose barb.
- 12. Clean all threads thoroughly to make sure nothing will bind them.
- 13. Insert motor assembly back into inner housing, **be careful not to damage O-rings.**
- 14. Connect control module wires back onto motor lead.

Note: A very small amount of deionized water may be used to help lubricate O-ring seals.

- 15. Twist the motor assembly and inner housing assembly 180 degrees counter-clockwise, coiling the wires to create more room.
- 16. Push the inner housing over the cap, **be careful not to damage O-rings.**
- 17. Install the outer housing making sure the threads do not bind.

REPLACEMENT PARTS



| REPLA | ACEMENT PARTS (not shown) | PART # |
|----------|------------------------------|----------|
| 1. GEO | SUB 2 CONTROLLER 300 W 120 V | 81200034 |
| 2. GEO | SUB 2 CONTROLLER 300 W 230 V | 81200035 |
| 3. MAN | UAL | 21200302 |
| 4. FEMA | ALE CONNECTOR KIT FOR CABLE | 51200307 |
| 5. MALE | E CONNECTOR KIT FOR MOTOR | 51200308 |
| 6. O-RII | NG KIT | 51200306 |
| 7. CHE | CK VALVE (OPTION) | 81200033 |
| 8. TAPE | GUIDE | 22050255 |
| 9. DRO | P TUBE | 51200299 |
| 10. DRC | OP TUBE INTAKE | 51150078 |

FOR MAINTENANCE AND SERVICE PLEASE CALL GEOTECH DIRECTLY AT 1-800-833-7958